

AD A 039642

39

RESEARCH AND DEVELOPMENT BRANCH
DEPARTMENT OF NATIONAL DEFENCE
CANADA

DEFENCE RESEARCH ESTABLISHMENT OTTAWA

DREO TECHNICAL NOTE NO. 77-2
DREO TN 77-2

UPDATING THE IM-5016 WIDE-RANGE GAMMA SURVEY METER

by

P.C. East and M.A. Periard



DISTRIBUTION STATEMENT A
Approved for public release
Distribution Unlimited

DDC
RECEIVED
APR 19 1977
C

AD No. _____
DDC FILE COPY

PROJECT NO.
16-03-01

APRIL 1977
OTTAWA

CAUTION

This information is furnished with the express understanding
that proprietary and patent rights will be protected.

DEPARTMENT OF NATIONAL DEFENCE
CANADA

TECHNICAL NOTE NO. 77-2

by
P.C. East and M.A. Periard
Nuclear Effects Section
NBC Defence Division

⑪ Apr 77

12) 12p.

(14) DREG-TN 77-2

⑩ Philip C. East Michael A. Periard

77-085

RECEIVED APRIL 1977
PUBLISHED APRIL 1977
OTTAWA

404576

Y/B

UNCLASSIFIED

ABSTRACT

This note describes a number of modifications to the IM-5016 low-range radiacmeter. The range has been extended to two three-decade scales, 0.1 to 100 mR/h and 0.1 to 100 R/h with automatic range changing. The modified instrument is powered by a single "D" size cell.

RESUME

Le présent rapport a pour but de présenter quelques modifications au radiamètre IM-5016. La gamme de mesure est plus étendue, ayant deux échelles de lecture de trois décades, soit de 0.1 à 100 mR/h et de 0.1 à 100 R/h. L'échelle est déterminée automatiquement par l'appareil et la lecture se fait à l'aide d'une diode photogène clignotante. L'appareil est alimenté par une pile 1.5 V de type "D".

INTRODUCTION

The IM-5016 is a low-range gamma survey meter that has been in service with the Canadian Forces for over 10 years. This instrument measures gamma dose rates from 1 mR/h to 10 R/h on two quasi-logarithmic scale ranges: 1 mR/h to 100 mR/h and 0.1 R/h to 10 R/h. A commercial version of this instrument is used by the Emergency Planning Canada and by industry in Canada and abroad. While this instrument has been successful it does have a number of limitations: -

1. The military version uses an expensive mercury battery originally required to meet the low-temperature requirement.
2. To cover a range of four decades mechanical range switching is used.
3. With two logarithmic scales there is the possibility of the operator introducing a gross error due to reading the wrong scale.
4. A mechanical offset in the meter zero, due to mechanical shock or temperature cycling, can introduce large scale errors.

A digital meter has been described (1) which overcomes these limitations. Unfortunately this instrument is relatively expensive and there are objections (2) to the use of a digital readout in a portable survey meter because of the difficulty in interpreting the readings when moving about in a non-uniform field.

This note describes a modification to the IM-5016 which: -

1. Converts the circuit to single "D" cell operation.
2. Changes the range automatically.
3. Increases each range to three decades, 0.1 mR/h to 100 mR/h or 0.1 R/h to 100 R/h, and aligns the scales so that only one set of gradations is required.
4. Adds a red light-emitting diode to the meter scale plate which flashes whenever the circuit switches to the high range.

These modifications should be commercially attractive for small-scale production and could be used to update existing IM-5016's. A photograph of a modified instrument is shown in Figure 1.




Figure 1. Photograph of Modified IM-5016

CIRCUIT DESCRIPTION

The circuit diagram is shown in Figure 2. For dose rates below 100 mR/h the circuit is on the low range and both G-M tubes are operating. The low-range tube, an 18506 has a nominal count rate of about 6 c/S at 1 mR/h. The G-M tube output pulses are fed via transistor switch Q1 to amplifier Q2, Q3. Transistor Q3 drives transistors Q4 to Q7. This is a variation of the diode pump circuit (3), so that condensers C1 to C4 are discharged by each G-M tube pulse. The charging current to C1, and C3 provides the meter current over the three decades of dose rate. The fourth stage, C4, produces a significant meter current only near full scale and provides compensation for the dead-time losses of the G-M tubes.

Range switching is controlled by flip-flop F1 and is normally set to the low range unless both F1 and F2 are clocked by gate G6. Gates G3, G5 and G6 are connected to counter C1 so that G6 will clock F1 and F2 after 112 pulses from the high-range G-M tube, an 18529. The 18529 has a nominal count rate of 35 c.s at 0.1 R/h or 112 counts in approximately 3 seconds. With F1 clocked: -

1. Transistor Q1 is off and the low-range G-M tube is off.
2. Transistor Q9 is off and Q8 is on connecting the high-range calibration potentiometer across the meter.
3. Gate G12 is on and the high-range LED indicator flashes.
4. Gate G5 is off so that F1 and F2 are clocked after 96 counts in each timing cycle or 86 mR/h.

With F2 clocked, gate G8 is off and the timing circuit cannot set F1 to the low range. If the dose rate falls below 86 mR/h, F2 will not be clocked and at the end of the timing cycle F1 is set to the low range.

The count rate of the high-range tube is about six times that of the low-range tube for the same scale reading. Gates G1 and G2 pass the first three of each four pulses from the Q2 output of counter C1, thus the output of G2 feeds the high-range G-M tube pulses divided by 16/3 to the pump circuit.

POWER SUPPLIES

The circuit diagram of the power supply is shown in Figure 3. The 9-volt supply must be well regulated since the meter current is directly proportional to it and a 1% change in the meter current will cause a 10%




Figure 2. Circuit Diagram




Figure 3. Converter Circuit

change in the scale reading on a three-decade logarithmic scale.

The circuit is regulated by a Zener diode fed by an essentially constant current and there will be negligible variation in the 9-volt supply over the useful life of the battery at room temperature. With change in temperature, the 9-volt supply will vary due to the temperature coefficient of the Zener diode. The base-emitter voltage variation of Q13 is compensated by transistor Q14, strapped as a diode.

The LED is fed directly from a tap on the collector winding and is on during the oscillator transistor Q12 off-time. For a constant output voltage and load, this off time is constant. To supply this load the collector current of Q12 and its on time must increase as the battery voltage drops. Thus the average current to the LED reduces as the battery voltage drops. For adequate brightness at low battery voltage, there is wasted power at higher voltages. This is preferred to the constant loss in a rectifier plus the cost of a diode and a condenser.

The zinc-carbon cell should provide over 40 hours of continuous operation at moderate temperature. To meet the low-temperature requirements the lithium cell must be used.

PERFORMANCE

The performance of this instrument will be similar to that of the IM-5016.

The detectors, the detector housing and location are unchanged, so that the energy dependence and directional characteristics will be as described for the IM-5016 (4).

The scale accuracy requirement of $\pm 20\%$ should easily be met over the four decades 1 mR/h to 10 R/h. For dose rates below 1 mR/h the accuracy will be limited by poor statistics and for dose rates above 10 R/h the errors may be aggravated by the difference in the dead-time errors of the two G-M tubes.

The circuit is relatively insensitive to temperature change. The meter circuit is arranged so that its change in sensitivity due to the $0.4\% ^\circ\text{C}$ temperature coefficient of the meter resistance compensates for the change in the 9-volt supply due to the temperature coefficient of the Zener diode, typically $0.6\% ^\circ\text{C}$. Temperature errors should be less than $\pm 10\%$.

REFERENCES

1. A Digital, Dual-Range, Portable Gamma Survey Meter Incorporating Dose Measurement. P.C. East and M.A. Periard. DREO TN 76-29
2. Private Communication. M. Walsh, Central Health Physics, Ontario Hydro. 1976
3. Transistor Circuit Design for a Radioactivity Contamination Meter. F.S. Goulding, Atomic Energy of Canada Ltd. Feb. 1957. AECL Report CREL-605
4. Production Approval Testing of Radiacmeter, Gamma Survey, Low Range, IM-5016/PD R.H. Panther, DAEL Feb. 1966

UNCLASSIFIED

Security Classification

DOCUMENT CONTROL DATA - R & D		
(Security classification of title, body of abstract and indexing annotation must be entered when the overall document is classified)		
1 ORIGINATING ACTIVITY Defence Research Establishment Ottawa National Defence Headquarters		2a DOCUMENT SECURITY CLASSIFICATION Unclassified
		2b GROUP N/A
3 DOCUMENT TITLE Ottawa, Ontario KIA 024 Updating the IM-5016 Wide Range Gamma Survey Meter (U)		
4. DESCRIPTIVE NOTES (Type of report and inclusive dates) Technical Note		
5 AUTHOR(S) (Last name, first name, middle initial) East, Philip C. and Periard, Michael A.		
6 DOCUMENT DATE March 1977	7a. TOTAL NO. OF PAGES 7	7b NO. OF REFS 4
8a. PROJECT OR GRANT NO. 16-03-01	9a ORIGINATOR'S DOCUMENT NUMBER(S) DREO T.N. 77-2	
8b CONTRACT NO.	9b. OTHER DOCUMENT NO.(S) (Any other numbers that may be assigned this document)	
10 DISTRIBUTION STATEMENT		
11 SUPPLEMENTARY NOTES	12. SPONSORING ACTIVITY DREO	
13. ABSTRACT <p>→ This note describes a number of modifications to the IM-5016 low-range radiacmeter. The range has been extended to two three-decade scales, 0.1 to 100 mR/h and 0.1 to 100 R/h with automatic range changing. The modified instrument is powered by a single ²D size cell. (Unclassified)</p> <p>↖</p>		

UNCLASSIFIED

Security Classification

KEY WORDS

Radiacmeter
Gamma Survey
Gamma Dose rate
Fallout Monitoring
Low Level Radiacmeter

INSTRUCTIONS

1. **ORIGINATING ACTIVITY** Enter the name and address of the organization issuing the document.
- 2a. **DOCUMENT SECURITY CLASSIFICATION** Enter the overall security classification of the document including special warning terms whenever applicable.
- 2b. **GROUP** Enter security reclassification group number. The three groups are defined in Appendix 'M' of the DRR Security Regulations.
3. **DOCUMENT TITLE** Enter the complete document title in all capital letters. Titles in all cases should be unclassified. If a sufficiently descriptive title cannot be selected without classification, show title classification with the usual one-capital letter abbreviation in parentheses immediately following the title.
4. **DESCRIPTIVE NOTES** Enter the category of document, e.g. technical report, technical note or technical letter. If appropriate, enter the type of document, e.g. interim, progress, summary, annual or final. Give the inclusive dates when a specific reporting period is covered.
5. **AUTHOR(S)** Enter the name(s) of author(s) as shown on or in the document. Enter last name, first name, middle initial. If military, show rank. The name of the principal author is an absolute minimum requirement.
6. **DOCUMENT DATE** Enter the date (month, year) of Establishment approval for publication of the document.
- 7a. **TOTAL NUMBER OF PAGES** The total page count should follow normal pagination procedures, i.e., enter the number of pages containing information.
- 7b. **NUMBER OF REFERENCES** Enter the total number of references cited in the document.
- 8a. **PROJECT OR GRANT NUMBER** If appropriate, enter the applicable research and development project or grant number under which the document was written.
- 8b. **CONTRACT NUMBER** If appropriate, enter the applicable number under which the document was written.
- 9a. **ORIGINATOR'S DOCUMENT NUMBER(S)** Enter the official document number by which the document will be identified and controlled by the originating activity. This number must be unique to this document.
- 9b. **OTHER DOCUMENT NUMBER(S)** If the document has been assigned any other document numbers (either by the originator or by the sponsor), also enter this number(s).
10. **DISTRIBUTION STATEMENT.** Enter any limitations on further dissemination of the document, other than those imposed by security classification, using standard statements such as:
 - (1) "Qualified requesters may obtain copies of this document from their defence documentation center."
 - (2) "Announcement and dissemination of this document is not authorized without prior approval from originating activity."
11. **SUPPLEMENTARY NOTES.** Use for additional explanatory notes.
12. **SPONSORING ACTIVITY** Enter the name of the departmental project office or laboratory sponsoring the research and development. Include address.
13. **ABSTRACT.** Enter an abstract giving a brief and factual summary of the document, even though it may also appear elsewhere in the body of the document itself. It is highly desirable that the abstract of classified documents be unclassified. Each paragraph of the abstract shall end with an indication of the security classification of the information in the paragraph (unless the document itself is unclassified) represented as (TS), (S), (C), (R), or (U).

The length of the abstract should be limited to 20 single-spaced standard typewritten lines, 7 1/2 inches long.
14. **KEY WORDS** Key words are technically meaningful terms or short phrases that characterize a document and could be helpful in cataloging the document. Key words should be selected so that no security classification is required. Identifiers, such as equipment model designation, trade name, military project code name, geographic location, may be used as key words but will be followed by an indication of technical context.